

11th April, 2017.

REF: GDC/GRM/OT/042/2016-17.

TO: ALL BIDDERS

Dear Sirs,

ADDENDUM 1: CLARIFICATION

RE: TENDER FOR SUPPLY OF WELL TESTING EQUIPMENT AND WORKSHOP CONSUMABLES - GDC/GRM/OT/042/2016-17.

In response to bidder's request for clarification, GDC wishes to provide the following clarification;

Bidder's Question 1

Schedule 1 (standard pressure dial gauges)

- What is the dial diameter?
- Is it a back or a bottom entry dial gauge?
- What are the thread type BSP or NPT?

GDC Response

Schedule 1 (standard pressure dial gauges)

- Dial diameter is **4** inches
- It is a **bottom** entry dial gauge
- Threads type are **BSP**

Bidder's Question 2

Schedule 1 (Wellhead geothermal circular pressure recorders)

- What is the chart recorder diameter required (8 Or 12 inches)
- Are the charts required to be weather proof?

GDC Response

Schedule 1 (Wellhead geothermal circular pressure recorders)

- Chart recorder diameter required is 12 inches.
- Charts required do not have to be weather proof.

NB/ All other requirements/instructions remain unchanged.

Yours Faithfully



DORIS KYAKA
MANAGER, SUPPLY CHAIN

